



No.: AIIMS/R/CS/TNE/2025/429/LPC/

Date: 10/11/2025

**(QUOTATION NOTICE)**

Request for E-Market Survey for submission of Price Quotation for “Supply of Multipara Monitor Accessories for Department of Trauma & Emergency at AIIMS, Raipur under GFR-2017, Rule-155.

**NOTICE INVITING PRICE QUOTATION UNDER GFR-2017, RULE-155**

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST/ relevant documents for above said. The quotation should be submitted to Medical College Building, Ground Floor, Gate No. 05, AIIMS, Raipur office of the Stores Officer up to **17/11/2025** before **3:00 pm**. The quotations will be opened on the same day at **03:30 pm**. detailed specifications of items are as under:

Sl. No	Item of Description	Qty.	Make/ Model	HSN Code	Unit Price in (₹)	GST	Unit Rate with GST	Total Price
01	6 Lead ECG Cable	10						
02	Spo2 Sensor for Adult Patient	10						
03	Adult NIBP 15 cm width Range 26-36 cm	10						
04	Adult NIBP Tubing	10						

**Note: The above accessories must be compatible with the existing installed Multipara Monitor. (Model: BSM-3763)**

**Terms and Condition**

1. Rate should be mentioned in words & figure both.
2. Taxes/GST, if any (Kindly mention in above table)
3. Delivery schedule: within 45 days from the date of issue of P.O. Delivery period:
4. No additional documents related to this NIQ will be entertained after opening of NIQ.
5. Price should be FOR Destination basis (i.e. Department of Trauma & Emergency AIIMS Raipur).
6. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
7. Quotation No/Name and due date of opening must be written on top of envelop.
8. GST rates applicable on your quoted item may please be confirmed. **HSN code for item should be clearly mentioned.**
9. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following **Anti Profiteering Clause of GST.**

**“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.**

10. The GST registration details may be furnished.
11. RTGS details required for payment purpose.
12. 100% payment against receipt and acceptance of material.
13. No part payment will be entertained.
14. Validity of offer should not be less than 90 days.
15. Supply, installation and commissioning will be done by firm (if applicable).
16. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
17. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.
18. The product should be of superior quality and highly standard.
19. L-1 of items will be calculated on the total value of all goods.

**Stores Officer,  
AIIMS, Raipur (C.G.)**

## **Technical Specification**

### **1. 6 Lead ECG Cable**

- It should be a 6 lead ECG Cable.
- It should have built in advance arrhythmia monitoring on all leads.
- It should protect against surges produced by Electro surgical and Defibrillator potentials
- Sample may be asked for demonstration purpose to verify the quality & above mentioned specifications.

### **2. Spo2 Sensor for Adult Patient**

- Adult/Paediatric SPO2 sensor should be able to connect with monitors in Pediatric Department of AIIMS Raipur.
- The SPO2 Sensor should work on blu-pro/FAST technology. (Documents to be enclosed with technical bid)
- It should work on the principle of 2 wavelength light absorption method.
- It should be capable to measure SPO2 in low perfusion. (White Papers confirming the low perfusion technology to be enclosed with technical bid)
- SPO2 sensor should be water washable for cleaning purpose.
- It should pass drop test from a height of upto 10 feet.
- It should have finger-clip design and should be usable on patients weighing more than 20Kgs.
- The Sensor should be provided with an extension cable.
- Sample may be asked for demonstration purpose to verify the quality & above mentioned specifications.

### **3. Adult NIBP 15 cm width Range 26-36 cm**

- NIBP Cuff should have width of 13 cm
- It should have range from 26-36 cm
- It should not be made of natural rubber or latex
- It should be reusable in nature having Velcro durability for atleast 2,800 cycles.
- It should be soft in touch
- Accuracy should be +/-3 mmHg
- Sample may be asked for demonstration purpose to verify the quality & above mentioned specifications.

### **4. Adult NIBP Tubing**

- NIBP hose pipe should be of double tubing pattern for faster inflation & Deflation
- The Hose pipe should not get damaged upon twisting.
- Sample may be asked for demonstration purpose to verify the quality & above mentioned specifications.