



**Date: - 11/01/2020**

**Corrigendum**

**विषय/Sub:** Corrigendum in Notice Inviting Tender for “Supply of **Individual Donor Nucleic Acid Testing (NAT) System**” for Transfusion Medicine & Blood Bank Department at AIIMS, Raipur

**संदर्भ/Ref:** Tender No. AIIMS/R/CS/BB/18/NAT/OT

In context to following corrigendum is issued against subject tender:

| <b>Sr. No</b> | <b>Page/Clause/ Point</b> | <b>Existing Sentence</b>   | <b>To be read as/ Addendum</b>   |
|---------------|---------------------------|--|--|
| 01            | Page 17,<br>Point No.(3)  | 3. Amplification process must not be inhibited by Heparin or any other anticoagulant     | Amplification process must not be inhibited by common anticoagulants used in Blood bags/Blood Bank.                            |
| 02            | Page 17,<br>Point No.(6)  | 6. Test procedure must be able to target & amplify two separate regions of HIV-1 genome. | NAT Equipment should be able to target & amplify the HIV-1 genomes as per the provisions of National institute of Biologicals. |
| 03            | Page 18,<br>Point No.(34) | 34. Allequipment's/components /accessories shall be of the latest version.               | All equipment's/components /accessories shall be of the latest version globally.   |

Note: All other terms & condition will be remain unchanged.

**Sr. Administrative Officer  
AIIMS Raipur (CG)**