



No:- AIIMS/R/CS/OPHTHAL/E-13605/25/Corri.

Date: 29/07/2025

Corrigendum

Tender ID.No.2025_IMSRP_863325_1.

With reference to above Tender ID No., the following amendment is issued:

Annexure-A

S.No.	Page no And Point no in the NIT	Existing Tender Specification	To be read as
1.	Title of the Tender	Procurement of Dual Pump Transversal Phaco Machine for Micro Incision Cataract Surgery.	Dual Segment / Dual Pump Torsional / Transversal Phacoemulsification System.
2.	Page no. 43, Preferable features. Point No. 4.	The System should be able to perform Vitrectomy: Anterior& Posterior Vitrectomy as well as different modes of anterior vitrectomy I/A Cut, Visco.	The System should be able to perform Anterior Vitrectomy with different modes i.e. I/A Cut, Cut I/A Mode.
3.	Page no. 43, Required Accessories. Point No. 5.	Phaco tip (2.2/2.8) titanium compatible with machines – 10	Phaco tip (2.2/2.8) titanium/SS compatible with machines – 10

All other terms and condition will remain unchanged.

**Stores Officer (C.S.),
AIIMS Raipur.**

Copy to:-

1. PS to Director, AIIMS, Raipur
2. DDA, AIIMS, Raipur
3. Medical Superintendent, AIIMS Raipur.
4. FA, AIIMS, Raipur.
5. HOD, Department of Ophthalmology, AIIMS Raipur.
6. Accounts Officer (IFD), AIIMS Raipur.
7. Guard File.

